

RII – CHEST FOR THORACIC AORTA ANEURYSM REPAIR GE LIGHTSPEED 16 / OPTIMA CT580 PROTOCOL

Indications: Evaluate patency of stent graft, to determine thrombosis of excluded portion of aorta, and to look for endovascular leaks.

Position/Landmark	Head first or feet first-Supine Sternal Notch			
Topogram Direction	Craniocaudal			
Respiratory Phase	Inspiration			
Scan Type	Helical			
KV / mA / Rotation time (sec) Pitch / Speed (mm/rotation) Noise Index / ASiR / Dose Reduction	120kv / smart mA (100-440) / 0.5 sec 1.375:1 , 27.50mm 19.0 / 30 / 30%			
Detector width x Rows = Beam Collimation	1.25mm x 16 = 20mm			
Average Tube Output	Each Helical: ctdi – 10.3 mGy dlp – 415 mGy.cm			
First Helical Set		body	thickness/ spacing	recon
Slice Thickness/ Spacing	recon	part	algorithm	destination .
Algorithm	1	non con chest	5mm x 5mm	pac
Recon Destination	2	lung	5mm x 5mm	pac
Second Helical Set		body	thickness/ spacing	recon
Slice Thickness/ Spacing	recon	part	algorithm	destination .
Algorithm	1	arterial chest	2.5mm x 2.5mm	pac
Recon Destination	2	thin chest	1.25mm x .6mm	for dmpr
Third Helical Set		body	thickness/ spacing	recon
Slice Thickness/ Spacing	recon	part	algorithm	destination .
Algorithm	1	delayed chest	2.5mm x 2.5mm	pac
Recon Destination	2	thin chest	1.25mm x .6mm	for dmpr
Scan Start / End Locations	1cm superior to lung apices mid kidney 38cm			
DFOV	decrease appropriately			
IV Contrast Volume / Type / Rate	100cc omni 350 / 4cc per second			
Scan Delay	arterial smart prep		delayed 120 seconds	
2D/3D Technique Used	DMPR of 5mm x 5mm coronal chest series of the arterial and delayed phases (auto-batch on), average mode, auto-transferred to PACS.			
Comments:	The smart prep threshold for the arterial phase is +100 hu at the proximal thoracic aorta. This protocol is a non contrast, then arterial phase, then delayed phase of the chest to assess thoracic aorta aneurysm repair.			
Images required in PACS	Scouts, 5mm x 5mm axial non con chest, 5mm x 5mm axial non con lungs, 2.5mm x 2.5mm axial arterial chest, 5mm x 5mm coronal arterial chest, 2.5mm x 2.5mm axial delayed chest, 5mm x 5mm coronal delayed chest, Dose Report			