

**RIH – PEDI SURGICAL/3D HEAD
SIEMENS DEFINITION AS20 PROTOCOL**

Indications: This ct is performed to for pre-surgical planning of cranio-facial reconstruction.

Position/Landmark	Supine head first or feet first 1cm superior to skull vertex																																								
Topogram Direction	Craniocaudal / Craniocaudal																																								
Respiratory Phase	Any																																								
Scan Type	Helical																																								
Ref kV/Ref mAs/Rotation time (sec) Pitch / Speed (mm/rotation) Safire Strength / Dose Optimization	Care kV 120 / Care Dose4D 210 / 0.5 sec .7:1 , 8.75mm 1 / 3																																								
Detector width x Rows = Beam Collimation	0.625mm x 20 = 12.5mm																																								
Average Tube Output	ctdi – 25.0 mGy dlp – 500 mGy.cm																																								
Helical Set Slice Thickness/ Spacing Algorithm Recon Destination	<table border="1"> <thead> <tr> <th></th> <th>body part</th> <th>thickness/ spacing</th> <th>algorithm</th> <th>recon destination .</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>thick helical brain/face</td> <td>5mm x 5mm</td> <td>J40f medium</td> <td></td> </tr> <tr> <td>2</td> <td>axial brain reformat</td> <td>5mm x 5mm</td> <td>J40f medium</td> <td>pac</td> </tr> <tr> <td>3</td> <td>coronal brain reformat</td> <td>5mm x 5mm</td> <td>J40f medium</td> <td>pac</td> </tr> <tr> <td>4</td> <td>1mm true axial face skull</td> <td>5mm x 5mm</td> <td>H60f sharp</td> <td>pac</td> </tr> <tr> <td>5</td> <td>1mm true coronal face skull</td> <td>5mm x 5mm</td> <td>H60f sharp</td> <td>pac</td> </tr> <tr> <td>6</td> <td>1mm true sagittal face skull</td> <td>5mm x 5mm</td> <td>H60f sharp</td> <td>pac</td> </tr> <tr> <td>7</td> <td>1mm straight axial face skull</td> <td>1mm x 1mm</td> <td>H60f sharp</td> <td>pac/terarecon</td> </tr> </tbody> </table>		body part	thickness/ spacing	algorithm	recon destination .	1	thick helical brain/face	5mm x 5mm	J40f medium		2	axial brain reformat	5mm x 5mm	J40f medium	pac	3	coronal brain reformat	5mm x 5mm	J40f medium	pac	4	1mm true axial face skull	5mm x 5mm	H60f sharp	pac	5	1mm true coronal face skull	5mm x 5mm	H60f sharp	pac	6	1mm true sagittal face skull	5mm x 5mm	H60f sharp	pac	7	1mm straight axial face skull	1mm x 1mm	H60f sharp	pac/terarecon
	body part	thickness/ spacing	algorithm	recon destination .																																					
1	thick helical brain/face	5mm x 5mm	J40f medium																																						
2	axial brain reformat	5mm x 5mm	J40f medium	pac																																					
3	coronal brain reformat	5mm x 5mm	J40f medium	pac																																					
4	1mm true axial face skull	5mm x 5mm	H60f sharp	pac																																					
5	1mm true coronal face skull	5mm x 5mm	H60f sharp	pac																																					
6	1mm true sagittal face skull	5mm x 5mm	H60f sharp	pac																																					
7	1mm straight axial face skull	1mm x 1mm	H60f sharp	pac/terarecon																																					
Scan Start / End Locations	1cm inferior to chin 1cm superior to skull vertex 25cm																																								
DFOV	decrease appropriately																																								
IV Contrast Volume / Type / Rate																																									
Scan Delay																																									
2D/3D Technique Used	5mm x 5mm axial and coronal brain reformats, standard algorithm in respect to the glabello-meatal plane (auto-batch off), average mode, auto transferred to PACS 1mm x 1mm axial, sagittal, and coronal face/skull reformats, bone algorithm , in respect to the skull floor plane (auto-batch off), average mode, auto transferred to PACS 3d head tumble and spin.																																								
Comments: Since this study is comprised of all mpr's, Recon 1 is used only to acquire data. Recons 2-6 are workstream 4d reformats for pac. Recon 7 is thin pre-op planning image data to terarecon.																																									
Do not alter the pitch setting of this protocol.																																									
Images required in PACS	Topograms , 5mm x 5mm axial brain, 5mm x 5mm coronal brain, 1mm x 1mm axial, sagittal, and coronal face/skull reformats, bone algorithm, 1mm x 1mm prosthetic implant planning data set, 3d head tumble and spin, Patient Protocol																																								